

AD-A172 877

TECHNICAL EVALUATION OF DOSIMETER CORPORATION OF
AMERICA MODEL 005 POCKET.. (U) AIR FORCE OCCUPATIONAL
AND ENVIRONMENTAL HEALTH LAB BROOKS AF.. J A CALDWELL
AUG 86 USAFDEHL-86-069RE0094HXA


1/1

UNCLASSIFIED

F/G 18/4

ML





2

USAFOEHL REPORT



86-069RE0094HXA

**TECHNICAL EVALUATION OF DOSIMETER CORPORATION
OF AMERICA MODEL 005 POCKET DOSIMETER
USAFOEHL, BROOKS AFB TX**

JOHN A. CALDWELL, II, TSGT, USAF

August 1986

Final Report

DTIC
ELECTE
OCT 20 1986
B

AD-A172 877

DTIC FILE COPY

Distribution is Unlimited: approved for public release

USAF Occupational and Environmental Health Laboratory
Aerospace Medical Division (AFSC)
Brooks Air Force Base, Texas 78235-5501

NOTICES

When Government drawings, specifications, or other data are used for any purpose other than a definitely related Government procurement operation, the Government thereby incurs no responsibility nor any obligation whatsoever. The fact that the Government may have formulated, or in any way supplied the said drawing, specifications, or other data, is not to be regarded by implication, or other-wise, as in any manner licensing the holder or any other person or corporation; or conveying any rights or permission to manufacture, use, or sell any patented invention that may in any way be related thereto.

The mention of trade names or commercial products in this publication is for illustration purposes and does not constitute endorsement or recommendation for use by the United States Air Force.

Do not return this copy. Retain or destroy.

Air Force installations may direct requests for copies of this report to: USAF Occupational and Environmental Health Laboratory (USAFOEHL) Library, Brooks AFB TX 78235-5501.

Other Government agencies and their contractors registered with the DTIC should direct requests for copies of this report to: Defense Technical Information Center (DTIC), Cameron Station, Alexandria VA 22314.

Non-Government agencies may purchase copies of this report from: National Technical Information Service (NTIS), 5285 Port Royal Road, Springfield VA 22161.

The Public Affairs Office has reviewed this report, and it is releasable to the National Technical Information Service, where it will be available to the general public, including foreign nations.

This report has been reviewed and is approved for publication.


JOHN J. COUGHLIN, Colonel, USAF, BSC
Commander

Prepared by:



JOHN A. CALDWELL II, TSgt, USAF
NCOIC, Instrumentation and Calibration
Branch

Reviewed by:



DAVID C. ADAMS, Capt, USAF, BSC
Chief, Instrumentation and
Calibration Branch



DAVID R. CASE, Lt Col, USAF, BSC
Chief, Radiation Services Division

REPORT DOCUMENTATION PAGE

1a. REPORT SECURITY CLASSIFICATION Unclassified		1b. RESTRICTIVE MARKINGS N/A	
2a. SECURITY CLASSIFICATION AUTHORITY N/A		3. DISTRIBUTION / AVAILABILITY OF REPORT Distribution is unlimited; approved for Public Release	
2b. DECLASSIFICATION / DOWNGRADING SCHEDULE N/A		5. MONITORING ORGANIZATION REPORT NUMBER(S)	
4. PERFORMING ORGANIZATION REPORT NUMBER(S) USAFOEHL Report 86-069 RE0094HXA			
6a. NAME OF PERFORMING ORGANIZATION USAF Occupational and Environmental Health Laboratory	6b. OFFICE SYMBOL (if applicable) RZ	7a. NAME OF MONITORING ORGANIZATION	
6c. ADDRESS (City, State, and ZIP Code) Brooks AFB TX 78235-5501		7b. ADDRESS (City, State, and ZIP Code)	
8a. NAME OF FUNDING / SPONSORING ORGANIZATION Same as 6a	8b. OFFICE SYMBOL (if applicable)	9. PROCUREMENT INSTRUMENT IDENTIFICATION NUMBER	
8c. ADDRESS (City, State, and ZIP Code)		10. SOURCE OF FUNDING NUMBERS	
		PROGRAM ELEMENT NO.	PROJECT NO.
		TASK NO.	WORK UNIT ACCESSION NO.
11. TITLE (Include Security Classification) Technical Evaluation of Dosimeter Corporation of America Model 005 Pocket Dosimeter, USAFOEHL, Brooks AFB TX (U)			
12. PERSONAL AUTHOR(S) TSgt John A. Caldwell, II, USAF			
13a. TYPE OF REPORT Final	13b. TIME COVERED FROM TO	14. DATE OF REPORT (Year, Month, Day) August 1986	15. PAGE COUNT 11
16. SUPPLEMENTARY NOTATION			
17. COSATI CODES		18. SUBJECT TERMS (Continue on reverse if necessary and identify by block number)	
FIELD	GROUP	USAFOEHL DCA 005 Temperature Humidity	
		Dosimeter	
19. ABSTRACT (Continue on reverse if necessary and identify by block number) A technical evaluation of the DCA Model 005 Pocket Dosimeter was accomplished on 18-19 Jun and 1 Jul 86 at Kelly and Brooks AFB to access the variability caused by temperature and humidity extremes. Discrepancies were identified and considerations recommended.			
20. DISTRIBUTION / AVAILABILITY OF ABSTRACT <input type="checkbox"/> UNCLASSIFIED/UNLIMITED <input checked="" type="checkbox"/> SAME AS RPT. <input type="checkbox"/> DTIC USERS		21. ABSTRACT SECURITY CLASSIFICATION Unclassified	
22a. NAME OF RESPONSIBLE INDIVIDUAL TSgt John A. Caldwell II		22b. TELEPHONE (Include Area Code) (512) 536-2613	22c. OFFICE SYMBOL USAFOEHL/RZE

Contents

	Page
DD Form 1473	i
I. BACKGROUND	1
II. SURVEY PERSONNEL	1
III. PERSONNEL CONTACTED	1
IV. TEMPERATURE AND HUMIDITY PROCEDURE	1
V. TEMPERATURE AND HUMIDITY METHODOLOGY	2
VI. TEMPERATURE AND HUMIDITY RESULTS AND CONSIDERATIONS	2
VII. RESPONSE CAPABILITY PROCEDURE	2
VIII. RESPONSE CAPABILITY RESULTS AND CONSIDERATIONS	2
Distribution List	8

DTIC
ELECTE
OCT 20 1986

Approved	
DTIC	
Date	
By	
DTIC	
Date	
By	
A-1	

✓

I. **BACKGROUND:** As requested by San Antonio Air Logistics Center (SA-ALC), a technical evaluation of the Dosimeter Corporation of America (DCA) model 005 pocket dosimeter was accomplished to (1) determine variability with changes in temperature and humidity, and (2) verify the response capability of the dosimeters rejected by PMELs and returned to SA-ALC. On 18-19 Jun and 1 Jul 86 USAFOEHL personnel performed the evaluation at Kelly AFB and Brooks AFB TX.

II. **SURVEY PERSONNEL** TSgt John A. Caldwell, II

III. **PERSONS CONTACTED**

Capt Michael Stroup	Project Engineer, NDI Equipment	925-6961
1Lt Gary Schultz	Project Engineer, Radiation Detection Equipment	925-6961
Mr Luciano Guerra	Equipment Specialist, Radiation Detection Equipment	925-8635

IV. **TEMPERATURE AND HUMIDITY PROCEDURE**

A. A Sperry 160 kVp x-ray tube, model SPX-160, was used to irradiate the dosimeters in an environmentally controlled chamber.

B. The dosimeters were evaluated under two environmental extremes; 100°F/100% RH, and 20°F/20%RH. Both conditions were accomplished at ambient barometric pressure.

C. The dose rate from the x-ray unit was determined with the MDH Model 1015, OEHL No. 1883, Serial No. 1615, calibrated 24 Jan 86 with an MDH Mammography probe, Model 10X5-6M, Serial No. 7087.

D. The x-ray beam was attenuated using 2.5 mm of aluminum to reduce the dose rate of the x-ray unit to acceptable levels. The dose rate was 0.505 R/min at 1 meter.

E. The dosimeters were subjected to an integrated exposure of 250 mR with the x-ray unit operating at 80 kVp, 1.7 mA exposure.

F. The dosimeters were centered in the beampath of the unit and allowed to stabilize in the chamber for 15 minutes prior to the first exposure.

G. The dosimeters to be evaluated were separated into three groups; New, Repackaged, and Returned. Five dosimeters from each group were selected at random for the evaluation.

H. All indicated exposures in Tables 1 and 2 were the average of three consecutive exposures.

I. All readings were multiplied by an Air Density Correction Factor (ADCF) to increase the accuracy. The ADCF was calculated as follows:

$$\frac{273.15 + \text{TEMP}}{295.15} \times \frac{760}{\text{BP}}$$

Where: TEMP = Temperature in degrees Celsius
BP = Barometric Pressure in mm Hg

V. TEMPERATURE AND HUMIDITY METHODOLOGY

- A. All exposures were made in free air.
- B. All dosimeters were zeroed, exposed and read in the vertical position.
- C. The MDH x-ray monitor was used to determine exposure time required for an integrated exposure of 250 mR. The correction factor is the number by which the readings on the dosimeter should be multiplied to get the actual exposure which the dosimeter (and the user) received.
- D. The indicated exposure is the dose reading from the dosimeter.
- E. The correction factor was obtained as follows:

$$\text{Correction Factor} = (\text{Actual Exposure}) / (\text{Indicated Exposure})$$

VI. TEMPERATURE AND HUMIDITY RESULTS AND CONCLUSIONS

The data indicate that the response of the model 005 Pocket Dosimeter to x-ray radiation is approximately 6% to 50% low depending on the environmental conditions. Consideration should be given to generating an Air Density Correction Factor (ADCF) for each base location according to their particular environmental factors at the time or by the season. This would help compensate for the wide variability of the dosimeter readings. Note that ADCF calculations are not required by T.O. 33K7-2-1-1.

VII. RESPONSE CAPABILITY PROCEDURE

- A. The evaluation of the returned D-005 dosimeters was accomplished IAW T.O. 33K7-2-1-1. There were two deviations from the T.O.; all measurements indicated in Tables 3 and 4 are the average of three consecutive exposures, and ADCF was applied to increase the accuracy.
- B. Twenty (59%) of the original 34 dosimeters satisfied leak test criteria and were used in the evaluation. The 14 dosimeters that were not evaluated were rejected for quality control acceptance reasons; namely, hair-lines that could not be located.

VIII. RESPONSE CAPABILITY RESULTS AND CONSIDERATIONS:

A. As determined by the data in Tables 3 and 4 and IAW T.O. 33K7-2-1-1, the acceptance level would increase from 60% to 75% by the application of the ADCF to each dosimeter correction factor.

B. The reason for the return of the 13 acceptable dosimeters to SA-ALC/MMIRAC is unknown.

C. If there are any questions concerning this report, please contact TSgt Caldwell, AV 240-2613, Comm 536-2613.

TABLE 1

Response of the DCA Model 005 Pocket Dosimeter to X-Ray

18 Jun 86

PARAMETERS: 80 kVp, 1.7 mA, 2.5 mm Al Total Filtration
 0.505 R/min Dose Rate
 Actual Dose: 250 mR, Exposure Time: 30 sec.

CALIBRATION STANDARD: MDH Model 1015, OEHL No. 1883, SN 1615
 Used with the Mammography Probe Model 10X5-6M,
 SN 7087

ENVIRONMENTAL CONDITIONS: Temperature: 100°F, Relative Humidity: 100%
 Barometric Pressure: 741.9 mm Hg
 ADCF: 1.079

	<u>DOSIMETER SN</u>	<u>INDICATED EXPOSURE (mR)</u>	<u>CF W/O ADCF</u>	<u>CF W/ ADCF</u>
GROUP I	50051	240	1.042	1.124
	50054	225	1.111	1.199
	50100	235	1.064	1.148
	60190	230	1.087	1.165
	60198	243	1.028	1.109
GROUP II	10006	240	1.042	1.124
	10019	255	0.980	1.057
	10033	243	1.029	1.110
	10062	233	1.073	1.158
	10074	236	1.059	1.143
GROUP III	60209	325	0.769	0.829
	60234 *	-	-	-
	80327	166	1.506	1.625
	80340 *	-	-	-
	80346	246	1.016	1.096

Each value listed is average of three exposures

Average Correction Factor W/O ADCF = 1.067

+ 7.8% Difference

Average Correction Factor W/ADCF = 1.145

* These dosimeters failed because the hairlines were not visible at the beginning of the exposure procedure nor throughout two or more exposures.

TABLE 2

Response of the DCA Model 005 Pocket Dosimeter to X-Ray

19 Jun 86

PARAMETERS: 80 kVp, 1.7 mA, 2.5 mm Aluminum Total Filtration
 0.505 mR/min Dose Rate
 Actual Dose: 250 mR, Exposure Time: 30 sec.

CALIBRATION STANDARDS: MDH Model 1015, OEHL No. 1883, SN 1615
 Used with the Mammography Probe Model 10X5-6M, SN 7087

ENVIRONMENTAL CONDITIONS: Temperature: 20°F, Relative Humidity: 20%
 Barometric Pressure: 741.9 mm Hg
 ADCF: 0.971

	<u>DOSIMETER SN</u>	<u>INDICATED EXPOSURE (mR)</u>	<u>CF W/O ADCF</u>	<u>CF W/ ADCF</u>
GROUP I	50051	175	1.429	1.388
	50054	192	1.302	1.264
	50100	188	1.339	1.300
	60190	183	1.366	1.326
	60198	168	1.488	1.445
GROUP II	10006	180	1.389	1.349
	10019	205	1.219	1.184
	10033	215	1.163	1.129
	10062	180	1.389	1.349
	10074	233	1.073	1.042
GROUP III	60209	208	1.202	1.167
	60234 *	-	-	-
	80327	76	3.289	3.194
	80340 *	-	-	-
	80346	233	1.073	1.042

Each value listed is the average of three exposures

Average Correction Factor W/O ADCF = 1.494

- 9.6% Difference

Average Correction Factor W/ADCF = 1.398

* These dosimeters failed because hairlines were not visible at the beginning of the exposure procedure nor throughout two or more exposures.

TABLE 3

Response of the DCA Model 005 Pocket Dosimeter to Cs-137

For 100 mR Exposure, 1 July 86

PARAMETERS: 130 Ci Cs-137 J.L. Shepperd Model 81-10 Irradiator
 Dose Rate: 9.9 R/Hr @ 2 meters
 Exposure Time: 100 mR = 0.606 minutes

CALIBRATION STANDARD: Performed by Manufacturer on 20 Mar 86 using an MDH
 Model 2025, SN 2127, with a 3 cc probe, SN 4126,
 Traceable to the National Bureau of Standards

ENVIRONMENTAL CONDITIONS: Temperature: 24°C, Barometric Pressure: 745.4 mmHg
 ADCF: 1.026

<u>DOSIMETER SN</u>	<u>REASON FOR RETURN*</u>	<u>INDICATED EXPOSURE (mR)</u>	<u>CF W/O ADCF</u>	<u>CF W/ ADCF</u>
30031	NHL	93	1.07	1.098
30423	OT	125	0.80	0.821
30437	OT	112	0.90	0.923
50021	OT	120	0.83	0.852
50075	MDR	123	0.81	0.831
60117	MDR	122	0.82	0.841
60201	NHL	120	0.83	0.852
60208	NHL	122	0.82	0.841
60209	OT	123	0.81	0.831
60245	UNK	117	0.86	0.882
80307	NHL	113	0.88	0.903
80315	NHL	117	0.86	0.882
80326	NHL	118	0.85	0.872
80327	NHL	118	0.85	0.872
80339	NHL	117	0.86	0.882
80342	NHL	100	1.00	1.026
80343	NHL	108	0.92	0.944
80346	NHL	108	0.92	0.944
80390	OT	120	0.83	0.852
80437	NHL	112	0.90	0.923

*Reasons for Return by PMELs:

NHL = No Hair Line

OT = Out of Tolerance

MDR = Maintenance Deficiency Report

UNK = Unknown Discrepancy

Average Correction Factor without ADCF = 0.871

+2.26% Difference

Average Correction Factor with ADCF = 0.894

TABLE 4

Response of the DCA Model 005 Pocket Dosimeter to Cs-137

For 400 mR Exposure, 1 July 86

PARAMETERS: 130 Ci Cs-137 J.L. Shepperd Model 81-10 Irradiator
 Dose Rate = 9.9 R/Hr @ 2 meters
 Exposure Time: 400 mR = 2.424 minutes

CALIBRATION STANDARD: Performed by Manufacturer on 20 Mar 86 using an MDH
 Model 2025, SN 2127, with a 3 cc probe, SN 4126,
 Traceable to the National Bureau of Standards

ENVIRONMENTAL CONDITIONS: Temperature: 24°C, Barometric Pressure: 745.4 mmHg,
 ADCF: 1.026

<u>DOSIMETER SERIAL No.</u>	<u>REASON FOR RETURN*</u>	<u>400mR AVE</u>	<u>CF W/O ADCF</u>	<u>CF W/ ADCF</u>
30031	NHL	362	1.11	1.139
30423	OT	478	0.84	0.862
30437	OT	480	0.83	0.852
50021	OT	482	0.83	0.852
50075	MDR	477	0.84	0.862
60117	MDR	495	0.81	0.831
60201	NHL	463	0.86	0.882
60208	NHL	475	0.84	0.862
60209	OT	488	0.82	0.841
60245	UNK	472	0.85	0.872
80307	NHL	447	0.90	0.923
80315	NHL	445	0.90	0.923
80326	NHL	450	0.89	0.913
80327	NHL	473	0.85	0.872
80339	NHL	463	0.86	0.882
80342	NHL	425	0.94	0.964
80343	NHL	447	0.90	0.923
80346	NHL	437	0.92	0.944
80390	OT	472	0.85	0.872
80437	NHL	440	0.91	0.934

*Reasons for Return from PMELs:

- NHL = No Hair Line
- OT = Out of Tolerance
- MDR = Maintenance Deficiency Report
- UNK = Unknown Discrepancy

Average Correction Factor without ADCF = 0.878

+2.2% Difference

Average Correction Factor with ADCF = 0.900

DISTRIBUTION LIST

	Copies
AFOMS/SGPR Brooks AFB TX 78235-5000	1
HQ AFISC/SGM Norton AFB CA 92409-7001	1
HQ AFSC/SGPB Andrews AFB DC 20334-5000	1
OL AD, USAFOEHL APO San Francisco 96274-5000	1
USAF Rgn Med Cen Wiesbaden/SGB APO New York 09220-5300	1
SA-ALC/MMIRAC Kelly AFB TX 78241-5000	1
HQ AFLC/SGB Wright-Patterson AFB OH 45433-5001	1
DTIC Cameron Station Alexandria VA 22314	2

END

1/1-56

DTIC